



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
-----------------	-------------	----------------------	---------------------	------------------

10/803,214

03/18/2004

Lance Middleton

1842-0028

5833

28078 7590 02/26/2008

MAGINOT, MOORE & BECK, LLP

CHASE TOWER

111 MONUMENT CIRCLE

SUITE 3250

INDIANAPOLIS, IN 46204

EXAMINER

BOUCHELLE, LAURA A

ART UNIT

PAPER NUMBER

3763

MAIL DATE

DELIVERY MODE

02/26/2008

PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary	Application No. 10/803,214	Applicant(s) MIDDLETON ET AL.	
	Examiner LAURA A. BOUCHELLE	Art Unit 3763	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 06 December 2007.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1,2,7,8,13 and 14 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1,2,7,8,13 and 14 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
2. ☐ Certified copies of the priority documents have been received in Application No. _____.
3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413) |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date. _____ |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Response to Amendment

Claim Rejections - 35 USC § 112

The amendment filed 12/6/07 has overcome the previous 112 rejection, and that rejection has therefore been withdrawn.

Claim Rejections - 35 USC § 102

1. The text of those sections of Title 35, U.S. Code not included in this action can be found in a prior Office action.
2. Claims 1, 2, 8, 13, 14 are rejected under 35 U.S.C. 102(b) as being anticipated by Beller (WO 9608227).
3. Beller discloses a syringe to syringe mixing apparatus comprising an elongate body 3, nozzles 33, 34, 35 having a mixing chamber 30 there between, a first syringe 1, and a second syringe 2 that are received in the elongate body. Beller discloses a flow modifying element 4. With regard to the limitation in claim 1 calling for an enlarged mixing chamber, the examiner interprets the Beller reference to disclose an enlarged mixing chamber 30. The inner diameter of this chamber is larger than the inner diameter of the receiving portions 36, 37. The enlarged mixing chamber is located between the nozzle 35 and the portion at the opposite end 36 of the passageway. Beller further discloses a second nozzle 4 and intermediate mixing chambers between nozzle 35 and 4, and nozzles 4 and 33. See Fig. 5.

Claim Rejections - 35 USC § 103

4. The text of those sections of Title 35, U.S. Code not included in this action can be found in a prior Office action.

5. Claim 7 is rejected under 35 U.S.C. 103(a) as being unpatentable over Beller.
6. Claim 7 differs from Beller in calling for the first flow area to be about five times greater than the second flow area. Where the only difference between the prior art and the claims was a recitation of relative dimensions of the claimed device and a device having the claimed relative dimensions would not perform differently than the prior art device, the claimed device was not patentably distinct from the prior art device. See MPEP 2144.04.

Response to Arguments

7. Applicant's arguments filed 12/6/07 have been fully considered but they are not persuasive.
8. Applicant argues that Beller does not disclose a mixing chamber between the at least one nozzle and the end portion at both ends of the elongate passageway. The examiner believes that Beller meets the limitations as claimed and points applicant to Fig. 3 of the Beller disclosure. Beller shows a mixing chamber 30 between said at least one nozzle 35 and said portion 36 at each of said opposite ends of said passageway. The claim language is ambiguous and as interpreted by the examiner Beller meets the claim limitation.

Conclusion

9. **THIS ACTION IS MADE FINAL.** Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire **THREE MONTHS** from the mailing date of this action. In the event a first reply is filed within **TWO MONTHS** of the mailing date of this final action and the advisory action is not mailed until after the end of the **THREE-MONTH** shortened statutory period, then the shortened statutory period

will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to LAURA A. BOUCHELLE whose telephone number is (571)272-2125. The examiner can normally be reached on Monday-Friday 8-4.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Nicholas Lucchesi can be reached on 517-272-4977. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/Nicholas D Lucchesi/
Supervisory Patent Examiner, Art Unit 3763

Laura A Bouchelle
Examiner
Art Unit 3763

/L. A. B./

Examiner, Art Unit 3763

Application/Control Number: 10/803,214
Art Unit: 3763

Page 5